

# EC-CONFORMITY DECLARATION

We hereby declare that "**Stifneck**" - Extrication Collar models **980100 Baby No-Neck, 980200 Pediatric, 980300 No-Neck, 980400 Short, 980500 Regular, and 980600 Tall** which are manufactured by:

Laerdal Medical (Suzhou) Co. Ltd.  
 No.19 Building, Huoju Rd.  
 Science & Technology Industry Park  
 Suzhou New District  
 215009, Suzhou  
 P.R. China

and manufactured for:

Laerdal Medical AS  
 P.O.Box 377  
 N-4002 Stavanger  
 Norway

are manufactured in conformity with the product described in Product Configuration Document and the Product Device Master Record Index and other information found in the technical file.

Compliance is claimed for the essential requirements in council directive **93/42/EEC Medical Device Directive** regarding the harmonisation of the member-states' regulations regarding medical devices.

The product is in risk class I, according to Rule II in Appendix IX of the directive.


For Laerdal Medical AS

*Craig A. Remson*  
 Name  
 Technical Product Manager

*15/09/03*  
 Date



Laerdal Medical AS  
 P.O.Box 377  
 N-4002 Stavanger  
 Telephone: 51511700  
 Fax: 51511788

REV	DESCRIPTION OF CHANGE	CHANGE REFERENCE	ORIGINATOR	VERIFIED	APPROVED
			DATE AND SIGN	DATE AND SIGN	DATE AND SIGN
			DATE AND SIGN	DATE AND SIGN	DATE AND SIGN
A	Release	LN 03-055	<i>15/09/03</i> CA	<i>15/09/03</i> K	<i>15/09/03</i> K
THE INFORMATION CONTAINED HEREIN IS PROPRIETARY TO LAERDAL AND SHALL NOT BE USED FOR ANY PURPOSE DETRIMENTAL TO OR BE REPRODUCED WITHOUT PERMISSION OF LAERDAL		<input checked="" type="checkbox"/> DMR <input type="checkbox"/> DHF			
REFERENCE	REPLACES	REV	TITLE		
	LMC: 12301000		<b>Stifneck EC- Conformity Declaration (CE Marking)</b>		
PRODUCT GROUP	TOOL	PART NO	DOCUMENT NO	PAGE	ATTACH REV
AWS	NA	980X00	PRO-RP01-0140	1 of 1	0 A