

MegaCode Kelly Advanced MegaCode Kelly



ENGLISH

Directions for Use



www.laerdal.com



Laerdal
helping save lives

Table of Contents

Introduction 1
 Cautions and Warnings (Intended Use) 1
 Items Included 2
 Unpack 3
 a) Attaching the Legs 3
 b) VitalSim Connections 3
 Getting Started 3
 a) Cricothyroid Membrane 3
 b) Pupil Replacement 3
 In Use 3
 a) Obstructed Airway 3
 b) Tension Pneumothorax Decompression 3
 c) IV Cannulation 4
 d) Intramuscular Injections 4
 e) Urinary Catheterization 4
 VitalSim Functionalities 4
 a) Cardiac Related Skills 4
 b) Blood Pressure Arm 5
 c) Heart, Breath and Bowel Sounds and Speech 5
 After Use 5
 a) Cricothyroid Membrane/ Neck Skin Collars 5
 b) IV Cannulation 5
 c) Intramuscular Injection 5
 d) Urinary Catheterization 5
 e) Pneumothorax 6
 i) Mid-Clavicular 6
 ii) Mid-Axillary 6
 f) Chest Drain 6
 Maintenance 6
 a) Cleaning 6
 b) Extend Cleanliness 6
 c) Attaching/ Removing Arms 7
 d) IV Arm Skin and Vein Replacement 7
 TroubleShooting 8
 Technical Data 8
 a) General 8
 b) Wireless Remote Control 8
 c) MegaCode Kelly Manikin 8
 Parts 8

Introduction

MegaCode Kelly is a full-body, lifelike manikin designed for the practice of advanced, difficult and obstructed airway scenarios and IV therapy. When used with the VitalSim Vital Signs Simulator; MegaCode Kelly allows cardiac defibrillation, pacing with or without capture and variable threshold, ECG interpretation of over 1100 rhythm variations, measurement of non-invasive blood pressure, and the auscultation and recognition of heart, breath and abdominal sounds. VitalSim also allows the use of the pre-recorded vocal sounds, and live voice through the use of a wireless microphone and logging and scenario functions.

The manufacturing quality of this manikin should provide many sessions of training when reasonable care and maintenance are practiced.

Laerdal Recommends
 Endotracheal Tube – Size 7.5
 Laryngeal Mask Airway – # 4
 Combitube- Large Adult or Trainer
 KING LT – # 4
 Needle or Catheter – Size 21-gauge or smaller

 **Cautions and Warnings - Latex**

Cautions and Warnings

- This product contains **Natural Rubber** latex which may cause allergic reactions when in contact with humans.
- There are electronic components mounted inside the airway head. The following techniques should not be performed on this manikin due to the inability to properly sanitize the airway if they are performed:
 - Mouth-to-mouth/Mouth-to-mask ventilation
 - Insertion of simulated vomit for suctioning
- Lubricate the oral and nasal airways liberally with the lubricant provided prior to inserting any instrument, tube or airway device into the airway. Additionally, instruments and tubes should also be lubricated prior to use.
- Do not allow the manikin's skin to come in direct contact with ink or photocopied paper, as this can permanently stain the skin. Avoid using colored plastic gloves when handling the manikin, as they may also cause discoloration.
- Care should be taken when palpating the pulses to not use excessive force as this will result in no pulse being felt. Only two individual pulses can be felt at the same time.
- If a training session involves the administration of fluids and/or drugs into the IV arm, empty the arm immediately following the training session.
- **Defibrillation**
- Only apply the defibrillator to a defibrillation post or zap plate which is properly mounted on the manikin's chest.
- Do not provide more than 2 x 360J defibrillator discharges per minute as an average over a period of time to prevent overheating.
- The manikin chest must be kept dry. Special attention should be taken when using IV Arm.
- Do not apply conductive gel or conductive defibrillation pads intended for patient use to prevent chest skin pitting.
- Do not use cables or connectors with visible damage.
- Observe all normal safety precautions for use of defibrillators.
- Defibrillation must be performed over the two defibrillation connectors only.

This device generates, uses and possibly radiates radio-frequency energy. If it is not installed and used in accordance with the instructions, it may cause harmful interference to radio communications. In that case, the user is encouraged to attempt correction of the interference by:

- Reorienting or relocating the receiving antenna.
- Increasing the distance between the device and receiver.
- Connecting the device to an outlet on a circuit different from that to which the receiver is connected.
- Consulting the dealer or an experienced radio/TV technician for help.

Electromagnetic radiation from other radio transmitters or other electronic equipment may cause noise in the head speaker. To eliminate this noise, move manikin away from the radiation source or turn the head speaker volume to zero.

Items Included:

MegaCode Kelly VitalSim Advanced

- (1) Direction for Use
- (1) Laerdal Global Warranty Booklet
- (1) Adult, Full-body Manikin
- (1) Articulating IV Training Arm
- (1) Blood Pressure Training Arm
- (1) Manikin Lubricant
- (1) MegaCode Kelly Eye Set
- (6) Neck Skin Collars
- (1) Cricothyroid Membrane Tape Roll
- (1) Set of Replacement Pneumothorax Bladders
 - (2) Mid-Clavicular Bladders
 - (2) Mid Axillary Bladders
- (6) Chest Drain Modules
- (1) 10 Foot Manikin Connector Cable
- (1) Set of Defibrillation Posts
- (1) Carry Case
- (1) Jacket
- (1) Pants

MegaCode Kelly VitalSim Basic

- (1) Direction for Use
- (1) Laerdal Global Warranty Booklet
- (1) Adult, Full-body Manikin
- (1) Articulating IV Training Arm
- (1) Manikin Lubricant
- (6) Neck Skin Collars
- (1) MegaCode Kelly Eye Set
- (1) Cricothyroid Membrane Tape Roll
- (1) 10 Foot Manikin Connector Cable
- (1) Set of Defibrillation Posts
- (1) Set of Replacement Pneumothorax Bladders
 - (2) Mid-Clavicular Bladders
 - (2) Mid Axillary Bladders
- (1) Carry Case
- (1) Jacket
- (1) Pants

Skills Taught:

Airway Management Skills

- Obstructed airway
- Endotracheal intubation
- Nasotracheal intubation
- Digital intubation
- Oropharyngeal airway insertion
- Nasopharyngeal airway insertion
- Bag Valve Mask
- Retrograde intubation
- Lightwand intubation
- Laryngeal Mask Airway insertion
- Laryngeal Tube Insertion
- Combitube insertion
- Trans-Tracheal Jet Ventilation
- Surgical cricothyrotomy
- Needle cricothyrotomy
- Suctioning techniques
- Stomach auscultation to verify proper airway positioning

Drug Administration

- IV insertion, infusion and bolus into peripheral veins of forearm, antecubital fossa and dorsum of the hand
- Subcutaneous and intramuscular injections

Tension Pneumothorax Decompression

- Bilateral mid-clavicular sites
- Mid-axillary site (Right)

Chest Tube Insertion (VitalSim Advanced version only.)

- Mid-axillary site (Left)

Cardiac Related Skills (VitalSim versions only.)

- 3 – 4 lead ECG with over 1100 rhythm variations available for interpretation.
- Pacing with variable threshold, with or without capture
- Defibrillation capabilities (25 -360 joules)
- Programmable scenario base algorithms for instructor control

Circulatory Skills (VitalSim Advanced version only.)

- Bilateral carotid pulse
- Measurement of Noninvasive Blood Pressure
 - Brachial and radial pulses
 - Auscultation or palpation of noninvasive blood pressure measurement
 - VitalSim allow values to be set for systolic, diastolic pressures, heart rate, auscultatory gap and volume

Sounds and Speech (VitalSim Advanced version only.)


- 13 different heart sounds synchronized with ECG
- 8 different lung sounds
- 5 bowel sounds
- 9 preprogrammed vocal sounds
- Live voice through the use of a wireless microphone

Logging and Scenario Functions (VitalSim versions only.)

- PC scenario editor for user defined and pre-programmed scenarios.
- Student logs can be stored and downloaded for post training review and debriefing.
- VitalSim Control Unit will hold up to ten (10) scenarios and twenty-five (25) student logs.

Unpack:

MegaCode Kelly is packaged without the legs attached. Remove the upper body and legs from the packaging and attach the legs to the torso.



To attach legs:

Step 1 Remove thigh injection pads (A), genitalia module (B) and bladder reservoir assembly (C)
 Step 2 Select and prepare hardware (Photo 1), pass spring and small washer over bolt (Photo 2) Step 3 To install pass bolt assembly through thigh, then pelvis opening from outside inward (Photo 3), select large fender washer and pass over bolt from inside manikin abdominal cavity. Screw wing nut onto bolt (Photo 4) Step 4 Replace bladder reservoir and genitalia module in pelvis, injection pad in thigh. (See Step 1)

NOTE: Tighten wing nut as desired to simulate leg Range of Motion.

VitalSim™ Connections (VitalSim capable manikins only)

1. Connect MegaCode Kelly VitalSim Advanced to VitalSim Unit (photo 1), via 37 pin cable located on lower right side of manikin.
2. Connect the clear tubing exiting the manikin's right side to the back (Advanced Version).
3. Attach blood pressure cuff to manikin's left arm (Advanced Version).
4. Attach clear pneumatic tubing from blood pressure cuff (photo 2) to corresponding clear pneumatic tubing exiting the manikin's left (Advanced Version).



Photo 1



Photo 2

Getting Started

Cricothyroid Membrane/ Neck Skin Collar:

1. Cut a two (2) inch strip of Cricothyroid Membrane Tape.
2. Adhere tape to edges of cricoid opening, covering opening with tape, to create actual membrane.
3. Lay a Neck Skin Collar into molded track around neck area of manikin.
4. Attach collar using Velcro® strips.



Fig. 2

Laerdal Recommends

A tight seal will enhance actual feel and sound of penetrating cricothyroid membrane, and facilitate full rise and fall of the chest during bag valve mask ventilation.

Pupil Replacement

MegaCode Kelly is delivered with a set of pupils containing normal pupils mounted in the head. A separate case contains 3 sets of plastic pupil inserts (normal, constricted and dilated pupils) for use in simulating other conditions.

To change the pupils in the airway head:

1. Open the eyelids wide, take care not to rip the faceskin.
2. Using the suction cup provided or the edge of your fingernail, remove the pupil from the eye.
3. Replace with the pupil of choice, using the small suction cup tool provided or press into place with finger.

In Use

Laerdal Recommends

- 7.5 Endotracheal tube
- # 4 LMA
- Large Adult or Trainer Combitube
- KING LT – # 4

Airway Management:

1. Spray inside of pharynx, nostrils and all intubation tubes with a liberal amount of provided manikin lubricant or liquid soap. Sit the manikin upright to allow lubricant to coat passages.
2. Perform further intubation procedures as per your local training protocol.

NOTE: Should incorrect ventilation be performed, air will pass through esophagus, causing distention of the abdomen.

Obstructed Airway (Manual Inflation of Tongue):

1. Locate inflation bulb at lower left side of torso labeled "tongue."
2. Tighten air release knob.
3. Pump bulb several times to inflate tongue to desired size.

To deflate tongue, loosen air release knob.

Tension Pneumothorax Decompression (Manual Inflation of Pneumothorax):

1. Locate inflation bulb which exits lower left side of manikin, labeled "pneumo."
2. Tighten air release knob so that as you pump the bulb, no air escapes.
3. Pump bulb several times to prepare three (3) sites for decompression.
4. Decompress the site, bilateral mid-clavicular or right side mid-axillary or per your local training protocol.

When the site is "decompressed," air can be heard/observed passing through catheter.

IV Cannulation:

Laerdal Recommends

Use a 21G or smaller catheter for IV cannulation to extend the life of your manikin's IV arm.

1. Attach an IV bag with tubing. Note: Make sure the clamp is closed.
2. Use a syringe to inject blood concentrate and 500cc of water into an IV bag with tubing. Note: Make sure the clamp is closed. This will serve as the "blood source" bag. (Photo 3)



Photo 3

3. Attach the "blood source" bag with IV tubing to one of the latex tubes exiting the manikin's right IV arm.
4. Attach an empty bag with IV tubing to the second latex tube exiting the manikin's right IV arm. This bag will serve as the collection reservoir.
5. Control flow of blood from arm, via clamp on collection reservoir tubing. Put collection bag on floor to allow the now closed system to gravity feed.
6. Hang "blood source" bag from IV pole and open clamp to allow concentrate to flow through arm. Note: A blood pressure cuff can be used as a pressure infuser. This will allow the blood bag to be placed on the floor or between the mattress and frame. (Photo 4)



Photo 4

7. When the "blood source" bag is empty, switch the "blood source" bag and reservoir: Hang the now full reservoir and place empty reservoir bag on floor to reverse flow of concentrate. NOTE: Be sure to adjust the flow regulator. The collection reservoir clamp should always be used to control flow rate.

The arm is now ready for cannulation. Train per local protocol.

NOTE: Do not use iodine as this will permanently stain the manikin

Laerdal Recommends

When using a catheter or other venipuncture device, spray with lubricant spray for easier insertion.

Intramuscular Injections:

Medications can be administered via intramuscular injections bilaterally in the deltoids and thighs and in the gluteal and ventro-gluteal locations.

Laerdal Recommends

Medications may be administered via local protocol. All precautions and safety measures must be followed during training.

Urinary Catheterization: (Genitalia Kit Optional)

1. Remove blank genitalia from the manikin.
2. Lift and rotate the pelvic pin to one side. NOTE: Failure to replace the pelvic pin will result in pelvic spread.
3. Lift the urinary reservoir.
4. Using the urinary valves provided, attach threaded end of the connector valve to the corresponding threaded urinary connector on the genitalia (Photo 5-7).
5. Place the C Clamp over the valve, leaving loose.
6. Slide the smooth end of valve the to meet the ridge on the connector.



Photo 5 (Slide tube to meet ridge on connector.)

7. Move the C Clamp over the connector and the valve.



Photo 6 (Area where clamp is placed.)

8. Squeeze to tighten the C Clamp. This will lock the C Clamp in place.



Photo 7 (Apply clamp and squeeze to tighten.)

9. Replace the genitalia assembly into the manikin torso.
10. **Replace Pelvic Pin** NOTE: Failure to replace the pelvic pin will result in pelvic spread.
11. Remove plug in urinary reservoir.
12. Using a 100cc syringe, fill reservoir with colored water.
13. Perform urinary catheterization as per your local training protocol.

VitalSim Manikin Functionalities:

(VitalSim capable manikins only.)

Cardiac Related Skills: (VitalSim capable manikins only)

MegaCode Kelly, when used with VitalSim, has over 1100 rhythms variations. External pacing with or without capture, defibrillation and synchronized cardioversion can also be performed.

NOTE: Prior to training, ensure the 37 pin connector cable exiting the manikin's right side is attached to the VitalSim Control Unit.

1. Compressions may be performed with one or two hands.
2. For monitoring purposes, lead placement snaps are located on manikin as follows:
 - i. RA (Right Arm)
 - ii. LA (Left Arm)
 - iii. LL (Left Leg)
 - iv. RL (Right Leg)

NOTE: If the ECG snaps/ defibrillation connectors do not attach to the clinical monitor/ defibrillator/ external pacer used by your institution, additional ECG snaps and or defibrillation/pacing cables are available for purchase. **Please contact your representative for correct adaptor.**

3. Two (2) Zap Plates (discs) and two (2) hands free adapters have been designed for use with defibrillators. Screw discs or hands free adaptors into post sockets located on apex and sternum of manikin. Rhythm may be read from manikin using a monitor by applying paddles to discs with firm pressure. **(Do not use gel or hands free self adhesive pads.)**

4. For manual defibrillation, place paddles on zap plates and press down firmly for good contact **NOTE:** Monophasic and Biphasic defibrillators may be used.

Laerdal Recommends

Caution: Follow defibrillation protocol by avoiding contact between the paddle and any of the electrode sites while defibrillating. A load of up to **360** joules can be delivered.

See **VitalSim DFU** for complete cardiac functionality and operating information.

Safety Precautions used during Defibrillation of a Patient/Trainer

1. Read and follow all safety and operating instructions provided with your defibrillator and associated equipment.
2. This trainer can be shocked with actual voltages and current used on a live patient. All precautions and safety measures must be used during the defibrillation and pacing phases of training. Failure to follow safety measures could result in injury or death to operators, students or onlookers not heeding these warnings.

Blood Pressure Arm – (VitalSim Advanced)

Left Arm Installation

The manikin is packaged with the left BP Arm attached (**ALS Simulator and MegaCode Kelly Advanced only**). The BP Arm is designed to rotate approximately 220°.

Caution: To prevent damage, do not over rotate the left BP Arm.



Photo 1



Photo 2



Photo 3



Photo 4 - Inside View Torso Left Side



Photo 5 - Outside View Left Shoulder Standard or Wound Arm

To detach the BP Arm:

STEP 1 Detach the chest skin, lift the chest plate and push it aside to

remove the stopper screw (photo 1) and release the BP Arm. **STEP 2** Take out the arm stopper, disconnect the BP cables and remove the BP Arm. **NOTE** Properly store the BP Arm and the attachment accessories for future use.

To install the Left Arm (standard or wound model):

The manikin includes the necessary arm hardware (photo 2)

STEP 1 Select the replacement arm, remove the deltoid pad **STEP 2** Insert the arm adapter (photo 3) located inside the torso on the left side **STEP 3** Pass the small washer over the bolt, thread the bolt and washer assembly through the openings of the arm shoulder, torso and arm adapter (photo 4 & 5) **Note:** Tighten the wing nut as desired to simulate Range of Motion **STEP 4** Replace the chest plate, the chest skin and the shoulder deltoid pad.

To re-attach the BP Arm, reverse the processes. Properly store the arm, the adapter and the hardware set for future use.

The blood pressure arm, when used with VitalSim, can be used to auscultate and palpate blood pressure. Korotkoff sound volume can be controlled, an auscultatory gap heard and pressure set in steps of 2 mmHg.

1. Place the supplied Blood Pressure cuff on Blood Pressure are (LEFT).
2. Attach clear tubing on cuff to matching clear pneumatic tubing exiting underneath the manikin's left arm. (Photo 8)



Photo 8

3. Ensure that the 37 pin connector cable and the clear pneumatic tubing, both exiting the manikin's right side is attached to the VitalSim Control Unit.

See **VitalSim DFU** for complete Blood Pressure Arm functionality and operating information.

Laerdal Recommends

Do not insert needles into blood pressure arm.

Heart, Breath and Bowel Sounds and Speech (VitalSim Advanced)

MegaCode Kelly when used with VitalSim, has heart, breath and bowel sounds, both normal and abnormal with adjustable instructor controlled volume. VitalSim also allows the use of the pre-recorded vocal sounds, and live voice through the use of a wireless microphone.

1. Ensure that the MegaCode Kelly VitalSim Advanced is connected to the VitalSim Control Unit, via 37 pin cable located on lower left side of manikin.

See **VitalSim DFU** for complete Heart, Breath and Bowel Sound Auscultation and Speech functionality and operating information.

After Use:

Cricothyroid Membrane/ Neck Skin Collar:

1. Remove neck skin collar, by detaching Velcro®.
2. Remove used Cricothyroid Membrane Tape from cricoid opening.
3. Cut a two (2) inch strip of Cricothyroid Membrane Tape.
4. Adhere tape to edges of cricoid opening, covering the opening with tape, to create the actual membrane.



Fig. 3

5. Lay a Neck Skin Collar into molded track around neck area of the manikin. (Fig 3) **NOTE:** If the used collar is still in good repair simply slide the collar in either direction for a fresh site. If the collar is showing wear, discard and replace with a new one.
6. Attach collar using Velcro® strips.

Laerdal Recommends

The neck skin collar is designed to provide multiple sites for needle and surgical techniques. When a fresh site is needed, slide the collar in either direction. If the neck skin collar is showing wear, discard and replace with a new collar.

IV Cannulation

After IV training is complete:

1. Close both clamps on IV tubing and detach bag tubing from arm tubing. Fill a 100cc syringe with water. Flush IV arm with syringe. Note: This should be done over a basin to catch any fluid.
2. Using the empty syringe push air through venous system to remove excess water.

Intramuscular Injections

After intramuscular injection training is complete:

1. Remove sponge from inside of injection pad skin. **NOTE: Do not store wet foam pads in skin. This will cause shrinkage of injection pad skin.**
2. Rinse sponge with water, allowing it to dry overnight.
3. Insert sponge back into injection pad skin. **NOTE:** A small amount of powder applied to sponge will ease insertion.
4. Replace the pad in the manikin.

Urinary Catheterization

After urinary catheterization is complete:

1. Remove complete genitalia assembly.
2. Working over a basin empty the reservoir.
3. Rinse thoroughly and allow to dry overnight.

Pneumothorax:

(A) Bilateral Mid-Clavicular Sites:

1. Detach the chest skin from torso from tabs at the shoulder and back and remove.
2. Remove the hard chest plate from the torso.
3. Disconnect bladder hose from Y connector, located on the underside of the chest plate. (see photo 9)



Photo 9

4. Pinch and remove the bladder through the opening between the 2nd and 3rd intercostal spaces on the top side of the hard chest plate. (See photo 10)



Photo 10

5. Trim bladder tubing to match original tubing length to fit this manikin style.
6. Insert the new pneumothorax bladder into the top side of the chest plate through the 2nd and 3rd intercostal space opening with the bladder tubing exiting through the back side of the chest plate. **NOTE:** The narrow edge of the bladder should be inserted closest to the sternum. (See photo 11)



Photo 11

7. Reconnect hose to Y hose connector.
8. Return chest plate to proper position on torso.
9. Replace chest skin over torso, being sure to secure skin at shoulders and back.
10. Fill puncture marks on the exterior of the chest skin with the Bladder Replacement Kit.

(B) Mid-Axillary Site (RIGHT):

1. Remove chest skin from tabs at the shoulder and back.
2. Remove pneumo bladder box from right side of torso. (Figure 4)




Fig. 4

3. Disconnect bladder hose from inline hose connector, being sure hose does not fall through hole and into torso.
4. Remove bladder from pneumo pad and discard.
5. Trim bladder tubing to match original tubing length to fit this manikin style. Connect new bladder hose to inline hose connector.
6. Fold and insert new pneumo bladder into pneumo box.
7. Replace chest skin over torso, being sure to secure skin at shoulder and back tabs.
8. Fill puncture marks on exterior of chest skin with the Bladder Replacement Kit.

NOTE: Bladders can be repaired using vulcanized or contact cement glue (not provided). Cover punctured surface of pneumothorax bladders with glue. Allow bladders to air dry completely overnight before replacing in the chest plate.

Chest Drain:

Mid Axillary Site (LEFT) (VitalSim Advanced)

1. Remove chest skin from tabs at shoulders and back.
2. Remove the chest tube drain module from the left mid-axillary site of the manikin. (See photo 12)



Photo 12

3. Insert a new chest drain module.
4. Replace the chest skin over the torso being sure to secure the skin at the shoulders and both sides.

Maintenance:

Cleaning

1. Clean with mild soap and water; do not submerge the manikin or parts in cleaning fluids or water.
2. Use only on clean surface. Avoid felt tipped markers, ink pens, acetone, iodine or other staining products and avoid placing the manikin on newsprint or inked lines of any kind.
3. To ensure longevity, each manikin should be cleaned after each training session and a general inspection should be conducted regularly.
4. Modules and all other parts should be drained and air-dried thoroughly before storage and disinfected when needed. After use of injection pads (use water only), accumulated water should be squeezed out. **Do not store wet foam pads in the skin.** To prevent mildew or mold, pads can be soaked in a mild solution of disinfectant and water or bleach and water. Squeeze excess solution from pads, allow them to dry, then store or reinsert in manikin.
5. Articulating parts will benefit from a light application of talcum powder prior to training sessions.

To extend the cleanliness of your manikin:

1. Use a small towel and lightly dust the manikin head and chest with a small amount of powder (provided). Wipe away any excess.
2. Use gloves when handling the manikin.
3. Avoid felt tipped markers, ink pens, iodine and newsprint, these will stain the manikin.

Laerdal Recommends

Use only on clean surface. Avoid felt tipped markers, ink pens, acetone, iodine or other staining products and avoid placing the manikin on newsprint or inked lines of any kind. These will stain the manikin.

Attaching/Removing arms:

1. Remove deltoid injection pad from upper arm.
2. Detach chest skin from tabs at shoulders and back.
3. Remove skin to reveal internal portion of upper chest.
4. Lift hard chest plate to reveal inside sockets for arm connections.
5. Place one washer on threaded 4 ½" bolt.
6. Insert threaded bolt through upper arm and through torso so that it is visible in chest cavity. (Figure 5)



Fig. 5

7. From inside of chest area, slip a washer, a spring and another washer on bolt.

8. Screw a wing nut on bolt and tighten until desired articulation is achieved.
9. Replace hard chest plate.
10. Pull chest skin back onto shoulder area.
11. Connect chest skin to torso at shoulder and back tabs.
12. Replace deltoid injection pad on upper arm.

To remove arms(s), reverse procedure.

IV Arm:

Replacing Skin and Veins:

Laerdal Recommends

When excessive leaking occurs at the puncture sites, either a new vein system or skin should be installed to reduce loss of fluid.

We recommend working at a basin when replacing skin and veins.

Replacing skin:

1. Cut off skin. This can be done with a sharp knife or scalpel. (Figure 6)
2. Discard skin.




Fig. 6

Replacing veins: (Keeping skin)

1. Lubricate inside of skin with liquid detergent; let it flow down into the finger area.
2. Begin at top of arm and slowly pull skin down and remove from arm. **Do not roll, as that will cause skin to bind.** Thumb will detach with skin.
3. Remove tubing from track in mandrel. Glue may need to be scraped away.
4. Rinse and dry vein grooves well and swab with alcohol. Be sure to remove any excess glue.




Fig. 7

5. Place new veins along grooves, (Figure 7) spot gluing as needed. (We recommend a fast-drying glue applied sparingly every 2 to 3 inches.)
6. Generously lather arm mandrel with liquid soap.

Laerdal Recommends

Heat arms skins before replacing on mandrel.
This can be done with a hair-dryer.

7. Slide hand into skin. (Photo 13)




Photo 13

8. Work skin over fingers, as with a glove. (Photo 14)



Photo 14

9. Work arm skin up, over mandrel. (Photos 15 & 16.)



Photo 15



Photo 16

Troubleshooting

If an error message appears on the remote control stating “no reception,” make sure that the manikin is turned on.

Technical Data

General

Operating Temperature:
 0°C - 40°C at 90% relative humidity, non-condensing
 Storage Temperature:
 -15°C - 50°C at 90% relative humidity, non-condensing

MegaCode Kelly Manikin

Blood pressure accuracy: +/- 2mm Hg
 Defibrillation: Average of 720 J /minute max

Emergency Cardiac Scenarios

- Synchronized variable heart rate, rhythm, abnormalities and duration
- Defibrillation (25 -360 J)

IV Arm

Accessible veins include median, basilic and cephalic

Standards/Approvals

CE The product, when carrying the CE-mark, is in compliance with essential CE requirements and other relevant provisions of council directive 1999/5/EC.

Replacement Parts:

<u>Product Number:</u>	<u>Description:</u>
312029.....	Kit, Skin/Vein Adult Multi-Venous IV Arm
380200.....	Arm Assembly, Blood Pressure
200-02350.....	Hardware Set, Arms To Torso
380405.....	Bladder Assembly, Mid Clavicular
380410.....	Post Set, ECG/Defib

380445.....	Shaft, Compression
380455.....	Skin, Chest
380460.....	Genitalia, Blank-Adult Male
380461.....	Kit, Female Genitalia W/Urinary Valves
380462.....	Kit, Male Genitalia W/Urinary Valves
380470.....	Belly plate Assembly
380471.....	Plug, Belly plate-Adult
380475.....	Pin, Pelvis
380600.....	Leg Assembly, Left Adult
380650.....	Leg Assembly, Right Adult
380700.....	Arm, Male Multi-Venous IV Training
381102.....	Strap Set, Head skin
381105.....	Neck skin Set (6)
381107.....	Teeth, Upper
381402.....	Pad Assembly, Deltoid Injection Adult
200-01850.....	Set, Bladder Mid Axillary
200-00250.....	Tape, Crico-Thyroid
200-00350.....	Case, Carrying-Full Body Adult Manikin
200-00550.....	Cuff Assembly, Blood Pressure VitalSim
200-00850.....	Chest foam, MegaCode Kelly VitalSim
200-01450.....	Cable/Tube Assembly, VitalSim to Manikin
200-01750.....	Thrust Assembly, MegaCode Kelly VitalSim
200-02050.....	Hardware Set, Legs to Pelvis W/ Tools
300-00150.....	Pad Assembly, Thigh Injection Adult
300-00250.....	Pad Assembly, Ventral Gluteal Injection Adult Male
300-00350.....	Pad Assembly, Gluteal Injection Adult Male
300-00450.....	Valve/Clamp Set, Urinary Hosp Adult
250-21050.....	Lubricant, Manikin-4 oz Bottle W/ Sprayer
300-00750.....	Blood, Simulated-4 oz Red
205-03650.....	Lung Assembly
200-02850.....	Skin, head w/Straps
200-03050.....	Set, Pupil Inserts
200-03150.....	Airway/Tongue Assy
200-04150.....	Chest Skin (Spacer Style)

Please contact Laerdal Customer Service for more information on Replacement Parts in other skin tones.